

2025 Quantitative Chemical (Ethanol) Proficiency Test

FTS-25-QUANT1 Summary Report

The Submission Deadline for this test was July 25, 2025

The test was manufactured by FTS at the FTS Laboratory Facility (127 Grand River Avenue, Williamston, MI 48895) and all activities were coordinated by Rebecca Smith (rsmith@forsci.com), Proficiency Test Program Manager. Ms. Smith is also authorizing the release of this report. This is the summary report issued on 8/14/25. No activities were subcontracted in regards to this test. FTS considers all reports confidential and does not release information regarding participant's results without authorization from that participant.

Summary

Test results were received in 74 of 89 tests distributed (83% response rate). Responses for UTICs p20252736 and p20252737 did not meet the requirements for statistical evaluation for Items 1 and 2, so these results were not utilized in the calculation of the test statistics. Responses for UTICs p20252702, p20252758, and p20252761 did not meet the requirements for statistical evaluation for Item 1 and will not be utilized in the statistics. Responses for UTICs p20252708 and p20252766 did not meet the requirements for statistical evaluation for Item 2 and will not be utilized in the statistics.

Summary statistics of participant's responses are as follows:

| | Item 1 (27%) | Item 2 (5%) |
|--|-----------------|-----------------|
| Mean | 27.65% | 4.97% |
| Median | 27.70% | 4.99% |
| Maximum | 29.10% | 5.20% |
| Minimum | 26.30% | 4.63% |
| Robust mean (H15) | 27.65% | 4.97% |
| Robust standard deviation (H15) | 0.52% | 0.12% |
| Assigned Value (X_{pt}) | 27.65% | 4.97% |
| Target standard deviation (σ_{pt}) | 1.94% | 0.35% |
| Expanded uncertainty of the assigned value ($U(X_{pt})$) | 0.16% | 0.04% |
| Mean of participant's reported measurement uncertainties ($k=2$) | 1.30% | 0.24% |
| Median of participant's reported measurement uncertainties ($k=2$) | 1.09% | 0.20% |
| Participants with $ z\text{-score} (z_i) < 2$ | 69 of 69 (100%) | 70 of 70 (100%) |
| Participants with $ z\text{-score} (z_i) > 2 \text{ & } \leq 3$ | 0 of 69 (0%) | 0 of 70 (0%) |
| Participants with $ z\text{-score} (z_i) > 3$ | 0 of 69 (0%) | 0 of 70 (0%) |
| Participants with $ E_n\text{-score} (E_n) \leq 1$ | 67 of 69 (97%) | 65 of 70 (93%) |
| Participants with $ E_n\text{-score} (E_n) > 1$ | 2 of 69 (3%) | 5 of 70 (7%) |

Robust statistics were calculated by utilizing an Excel Add-In provided by the Royal Society of Chemistry (<http://www.rsc.org/membership/networking/interestgroups/analytical/amc/software/RobustStatistics.asp>). The statistical methods, participant's responses, z-scores and E_n -scores are summarized below.

Statistical Methods

Invalid results or Extreme Outliers

Results that are identifiably invalid (e.g. expressed in the wrong units, transpositions and other gross errors) or are extreme outliers (e.g. outside the range of 3Sd of the mean of reported values) may be excluded before any statistical calculations.

Assigned Value

For this proficiency test, assigned value is calculated as the robust mean using the procedure described in “ISO 13528:2022, Statistical methods for use in proficiency testing by interlaboratory comparisons – Annex C”.

Between-Laboratory Coefficient of Variation

The between-laboratory coefficient of variation is a measure of the between laboratory variation that in the judgment of the study coordinator would be expected from participants given the analyte concentration. It is important to note this is not the coefficient of variation of participant results. For the purpose of this proficiency test, this figure will be 7% for Item 1 and Item 2.

Target Standard Deviation

The target standard deviation (σ_{pt}) is the product of the assigned value (X_{pt}) and the between laboratory coefficient of variation (CV) as presented in Equation 1. This value is used for calculation of participant z-scores.

$$\text{Equation 1: } \sigma_{pt} = X_{pt} * CV$$

z-Score

For each participant result a z-score is calculated according to Equation 2 below:

$$\text{Equation 2: } Z_i = \frac{(X_i - X_{pt})}{\sigma_{pt}}$$

where:

Z_i = z-score

X_i = participant result

X_{pt} = the study assigned value

σ_{pt} = the target standard deviation from *Equation 1*

E_n -score

The E_n -score is complementary to the z-score in assessment of laboratory performance.

E_n -score includes measurement uncertainty and is calculated according to Equation 3 below:

Equation 3:

$$(E_n)_i = \frac{(X_i - X_{pt})}{\sqrt{U(X_i)^2 + U(X_{pt})^2}}$$

where:

$(E_n)_i$ = E_n -score

X_i = participant result

X_{pt} = the study assigned value

$U(X_i)$ = the expanded uncertainty of the participant's result

$U(X_{pt})$ = the expanded uncertainty of the assigned value

The consensus of participants' results is not traceable to any external reference, so although expressed in SI units, metrological traceability has not been established.

The uncertainty is estimated as:

$$u(X_{pt}) = 1.25 \times S^* / \sqrt{p}$$

where:

$u(X_{pt})$ = robust mean standard uncertainty

S^* = robust mean standard deviation

p = number of results

The expanded uncertainty ($U(X_{pt})$) is the standard uncertainty multiplied by a coverage factor of 2, a level of confidence of approximately 95%.

Criteria for the evaluation of performance of participants

Evaluation of participant's performance is the responsibility of the subscriber and the accreditation body (if applicable) based on comparison of the participant's result to the consensus responses. FTS does not evaluate participant's performance. Typically, z-scores and E_n -scores are evaluated using the following criteria:

z-score with absolute value ($|z_i|$):

$|z_i| \leq 2$ is satisfactory;

$|z_i| > 2 \leq 3$ is questionable;

$|z_i| > 3$ is unsatisfactory.

E_n -score with absolute value ($|(E_n)_i|$):

$|(E_n)_i| \leq 1$ is satisfactory;

$|(E_n)_i| > 1$ is unsatisfactory.

Again, FTS does not perform this evaluation. For the convenience of subscribers, Responses with $|z_i|$ of greater than 2 or $|(E_n)_i|$ greater than 1 are highlighted in the summary report.

In general terms, the z-score provides a measure of how close the reported value is to the assigned value, with no relation to the reported uncertainty. The E_n -score provides a measure of how close the reported value is to the assigned value, considering the reported uncertainty. A participant that has a satisfactory z-score but an unsatisfactory En-score should evaluate whether their reported measurement uncertainty is sufficient.


Summary of Participant's Responses: Formatted for Statistical Analysis (Uncertainties Recalculated to k=2):

| UTIC | Webcode | Item 1 reported % | Item 1 Uncert. (% @ k=2) | Item 1: Z-score (zi) | Item 1: En score (En)i | Item 2 reported % | Item 2 Uncert. (% @ k=2) | Item 2: Z-score (zi) | Item 2: En score (En)i |
|-----------|---------|-------------------|--------------------------|----------------------|-------------------------|-------------------|--------------------------|----------------------|-------------------------|
| p20252701 | W049 | 27.80 | 1.20 | 0.08 | 0.12 | 5.00 | 0.60 | 0.09 | 0.05 |
| p20252702 | W163 | N/A | N/A | N/A | N/A | 5.10 | 0.00 | 0.37 | 2.91 |
| p20252703 | W088 | 28.10 | 1.09 | 0.23 | 0.41 | 5.09 | 0.19 | 0.34 | 0.60 |
| p20252704 | W088 | 27.70 | 1.09 | 0.03 | 0.05 | 5.01 | 0.19 | 0.11 | 0.20 |
| p20252705 | W088 | 28.00 | 1.09 | 0.18 | 0.32 | 5.00 | 0.19 | 0.09 | 0.15 |
| p20252706 | W088 | 27.70 | 1.09 | 0.03 | 0.05 | 4.99 | 0.19 | 0.06 | 0.10 |
| p20252707 | W088 | 27.30 | 1.09 | -0.18 | 0.32 | 5.09 | 0.19 | 0.34 | 0.60 |
| p20252708 | W088 | 27.70 | 1.09 | 0.03 | 0.05 | N/A | N/A | N/A | N/A |
| p20252709 | W088 | 27.60 | 1.09 | -0.03 | 0.05 | 4.89 | 0.19 | -0.23 | 0.42 |
| p20252710 | W088 | 27.90 | 1.09 | 0.13 | 0.23 | 5.04 | 0.19 | 0.20 | 0.35 |
| p20252711 | W088 | 28.30 | 1.09 | 0.34 | 0.59 | 5.08 | 0.19 | 0.31 | 0.55 |
| p20252712 | W088 | 28.10 | 1.09 | 0.23 | 0.41 | 4.95 | 0.19 | -0.06 | 0.10 |
| p20252713 | W088 | 27.50 | 1.09 | -0.08 | 0.14 | 4.95 | 0.19 | -0.06 | 0.10 |
| p20252714 | W088 | 27.20 | 1.09 | -0.23 | 0.41 | 4.85 | 0.19 | -0.34 | 0.63 |
| p20252715 | W088 | 27.90 | 1.09 | 0.13 | 0.23 | 4.98 | 0.19 | 0.03 | 0.05 |
| p20252716 | W088 | 26.90 | 1.01 | -0.39 | 0.73 | 4.85 | 0.19 | -0.34 | 0.63 |
| p20252717 | W088 | 27.30 | 1.09 | -0.18 | 0.32 | 4.94 | 0.19 | -0.09 | 0.15 |
| p20252718 | W088 | 27.40 | 1.09 | -0.13 | 0.23 | 4.89 | 0.19 | -0.23 | 0.42 |
| p20252719 | W088 | 27.70 | 1.09 | 0.03 | 0.05 | 4.95 | 0.19 | -0.06 | 0.10 |
| p20252720 | W088 | 27.20 | 1.09 | -0.23 | 0.41 | 4.90 | 0.19 | -0.20 | 0.35 |
| p20252721 | W088 | 27.10 | 1.09 | -0.28 | 0.50 | 4.63 | 0.18 | -0.97 | 1.85 |
| p20252722 | W088 | 28.50 | 1.09 | 0.44 | 0.77 | 5.13 | 0.20 | 0.46 | 0.77 |
| p20252723 | W088 | 27.80 | 1.09 | 0.08 | 0.14 | 5.05 | 0.19 | 0.23 | 0.40 |
| p20252724 | W088 | 27.80 | 1.09 | 0.08 | 0.14 | 5.00 | 0.19 | 0.09 | 0.15 |
| p20252725 | W088 | 27.50 | 1.09 | -0.08 | 0.14 | 5.06 | 0.19 | 0.26 | 0.45 |
| p20252726 | W088 | 27.50 | 1.09 | -0.08 | 0.14 | 4.98 | 0.19 | 0.03 | 0.05 |
| p20252727 | W088 | 27.90 | 1.09 | 0.13 | 0.23 | 5.07 | 0.19 | 0.29 | 0.50 |
| p20252728 | W088 | 28.30 | 1.09 | 0.34 | 0.59 | 5.16 | 0.20 | 0.54 | 0.92 |
| p20252733 | W092 | 27.20 | 1.10 | -0.23 | 0.40 | 5.00 | 0.20 | 0.09 | 0.15 |
| p20252735 | W031 | 26.87 | 1.67 | -0.40 | 0.46 | 4.89 | 0.31 | -0.23 | 0.26 |
| p20252738 | W044 | 26.30 | 2.27 | -0.70 | 0.59 | 4.97 | 0.43 | 0.00 | 0.00 |
| p20252739 | W114 | 27.30 | 1.47 | -0.18 | 0.24 | 5.00 | 0.33 | 0.09 | 0.09 |
| p20252740 | W114 | 27.60 | 1.47 | -0.03 | 0.03 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252741 | W114 | 29.10 | 1.53 | 0.75 | 0.94 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252742 | W114 | 28.00 | 1.47 | 0.18 | 0.24 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252743 | W114 | 28.70 | 1.53 | 0.54 | 0.68 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252744 | W114 | 27.30 | 1.47 | -0.18 | 0.24 | 4.80 | 0.27 | -0.49 | 0.63 |

| UTIC | Webcode | Item 1 reported % | Item 1 Uncert. (% @ k=2) | Item 1: Z-score (zi) | Item 1: En score (En)i | Item 2 reported % | Item 2 Uncert. (% @ k=2) | Item 2: Z-score (zi) | Item 2: En score (En)i |
|-----------|---------|-------------------|--------------------------|----------------------|-------------------------|-------------------|--------------------------|----------------------|-------------------------|
| p20252745 | W114 | 27.80 | 1.47 | 0.08 | 0.10 | 4.80 | 0.27 | -0.49 | 0.63 |
| p20252746 | W114 | 27.30 | 1.47 | -0.18 | 0.24 | 5.00 | 0.33 | 0.09 | 0.09 |
| p20252747 | W114 | 27.70 | 1.47 | 0.03 | 0.03 | 4.80 | 0.27 | -0.49 | 0.63 |
| p20252749 | W190 | 27.70 | 1.47 | 0.03 | 0.03 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252750 | W190 | 27.10 | 1.47 | -0.28 | 0.37 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252751 | W190 | 27.70 | 1.47 | 0.03 | 0.03 | 5.10 | 0.33 | 0.37 | 0.39 |
| p20252752 | W190 | 27.30 | 1.47 | -0.18 | 0.24 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252753 | W190 | 28.20 | 1.47 | 0.28 | 0.37 | 5.00 | 0.40 | 0.09 | 0.07 |
| p20252754 | W190 | 27.20 | 1.47 | -0.23 | 0.30 | 4.90 | 0.27 | -0.20 | 0.26 |
| p20252755 | W190 | 27.10 | 1.47 | -0.28 | 0.37 | 4.80 | 0.27 | -0.49 | 0.63 |
| p20252756 | W190 | 27.30 | 1.47 | -0.18 | 0.24 | 4.80 | 0.27 | -0.49 | 0.63 |
| p20252757 | W190 | 27.20 | 1.47 | -0.23 | 0.30 | 5.00 | 0.33 | 0.09 | 0.09 |
| p20252758 | W002 | N/A | N/A | N/A | N/A | 4.90 | 0.20 | -0.20 | 0.34 |
| p20252761 | W110 | N/A | N/A | N/A | N/A | 5.20 | 0.27 | 0.66 | 0.85 |
| p20252762 | W110 | 28.00 | 1.33 | 0.18 | 0.26 | 5.00 | 0.27 | 0.09 | 0.11 |
| p20252766 | W110 | 27.90 | 1.33 | 0.13 | 0.19 | N/A | N/A | N/A | N/A |
| p20252769 | W110 | 26.50 | 1.27 | -0.59 | 0.90 | 4.80 | 0.27 | -0.49 | 0.63 |
| p20252770 | W082 | 27.50 | 2.00 | -0.08 | 0.07 | 5.10 | 0.40 | 0.37 | 0.32 |
| p20252771 | W172 | 27.00 | 3.00 | -0.34 | 0.22 | 4.70 | 0.50 | -0.77 | 0.54 |
| p20252772 | W096 | 27.80 | 0.98 | 0.08 | 0.15 | 5.05 | 0.17 | 0.23 | 0.47 |
| p20252775 | W206 | 26.30 | 2.10 | -0.70 | 0.64 | 4.70 | 0.30 | -0.77 | 0.89 |
| p20252776 | W116 | 27.50 | 1.73 | -0.08 | 0.09 | 5.10 | 0.33 | 0.37 | 0.39 |
| p20252777 | W116 | 28.30 | 1.80 | 0.34 | 0.36 | 4.90 | 0.33 | -0.20 | 0.21 |
| p20252778 | W025 | 26.70 | 0.20 | -0.49 | 3.70 | 4.70 | 0.07 | -0.77 | 3.36 |
| p20252779 | W256 | 28.20 | 0.00 | 0.28 | 3.41 | 5.00 | 0.00 | 0.09 | 0.67 |
| p20252780 | W070 | 28.10 | 0.93 | 0.23 | 0.48 | 5.10 | 0.20 | 0.37 | 0.63 |
| p20252781 | W070 | 27.60 | 0.87 | -0.03 | 0.06 | 5.10 | 0.13 | 0.37 | 0.92 |
| p20252782 | W070 | 28.20 | 0.93 | 0.28 | 0.58 | 5.00 | 0.20 | 0.09 | 0.15 |
| p20252783 | W070 | 27.40 | 0.87 | -0.13 | 0.28 | 5.00 | 0.13 | 0.09 | 0.21 |
| p20252784 | W055 | 28.70 | 1.90 | 0.54 | 0.55 | 4.99 | 0.33 | 0.06 | 0.06 |
| p20252785 | W055 | 27.40 | 1.80 | -0.13 | 0.14 | 5.01 | 0.33 | 0.11 | 0.12 |
| p20252786 | W055 | 29.10 | 1.90 | 0.75 | 0.76 | 4.88 | 0.32 | -0.26 | 0.28 |
| p20252787 | W055 | 28.70 | 1.90 | 0.54 | 0.55 | 5.03 | 0.33 | 0.17 | 0.18 |
| p20252788 | W004 | 28.00 | 1.00 | 0.18 | 0.35 | 5.10 | 0.10 | 0.37 | 1.19 |
| p20252789 | W037 | 27.94 | 1.09 | 0.15 | 0.26 | 5.19 | 0.20 | 0.63 | 1.07 |

Item 1: Ethanol (27%)- Graph 1

Assigned Value =27.65% +/- 0.16% (k=2)



Error bars represent participant's stated uncertainty (k=2).


Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Item 1: Ethanol (27%)- Graph 2

Assigned Value = 27.65% +/- 0.16% (k=2)



Error bars represent participant's stated uncertainty (k=2).


Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Item 1: Ethanol (27%)- Graph 3

Assigned Value = 27.65% +/- 0.16% (k=2)



Error bars represent participant's stated uncertainty (k=2).


Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Item 2: Ethanol (5%)- Graph 1

Assigned Value = **4.97% +/- 0.04% (k=2)**



Error bars represent participant's stated uncertainty (k=2).


Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Item 2: Ethanol (5%)- Graph 2

Assigned Value = **4.97% +/- 0.04% (k=2)**



Error bars represent participant's stated uncertainty (k=2).


Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Item 2: Ethanol (5%)- Graph 3

Assigned Value = **4.97% +/- 0.04% (k=2)**



Error bars represent participant's stated uncertainty (k=2).

Independent estimates of analyte concentration with associated uncertainties (k=2):

Md = Robust Median

R.A. = Robust Average

Manufacturer's Information

An ethanol stock solution was prepared by weighing 0.0370 gram of Anhydrous Caffeine (Powder City, Lot #CAFAH20160215) on a calibrated Mettler AE163 balance. The caffeine was added to the contents of a sealed 500ml bottle of 200 proof Ethyl Alcohol (Calbiochem OmniPur, Batch 61271), mixed, and allowed to dissolve completely. This stock solution was utilized for Item 1. The caffeine was added for internal QA/QC purposes.

A second ethanol stock solution was prepared by weighing 0.0374 gram of Anhydrous Caffeine (Powder City, Lot #CAFAH20160215) on a calibrated Mettler AE163 balance. The caffeine was added to the contents of a sealed 500ml bottle of 200 proof Ethyl Alcohol (Sigma Aldrich, Batch SHBS1883), mixed, and allowed to dissolve completely. This stock solution was utilized for Item 2. The caffeine was added for internal QA/QC purposes.

Item 1 (~27% solution) was prepared by combining 486mL of the Ethyl Alcohol stock solution to 1314mL distilled water. The volumes were measured using Class A volumetric flasks and pipettes. The solution was mixed and packaged into Qorpak 4 dram Glass Vials (15 mL), (12112015144, GLC-00993; Lot 01092025144, GLC-00993).

Item 2 (~5% solution) was prepared by combining 90mL of the Ethyl Alcohol stock solution to 1710mL distilled water. The volumes were measured using Class A volumetric flasks and pipettes. The solution was mixed and packaged into Qorpak 4 dram Glass Vials (15 mL), (Lot 01092025144, GLC-00993).

After homogeneity testing, the glass vials were labeled with a UTIC number, taped around the cap with packaging tape, and wrapped in tri-fold paper towel. The Items were packaged into a 125 mL Wide Mouth Sample Vial (Berlin Packaging, Item 1181126) and labeled per FTS guidelines.

Both items with matching UTICs were packaged together in a cardboard box, sealed and labeled per FTS guidelines.

Homogeneity & Stability Quality Control Testing

Homogeneity testing was performed after the samples were packaged in glass vials, hand labeled in fill order. Ten of the vials for Items 1 and 2 were selected by random sampling throughout the fill sequence by selection of the hand labeled fill number with computer random number generation software (Ablebits Random Number Generator, Add-in Express, Ltd., Belarus, BY).

Each selected sample was homogenized, and then two approximately 5 ml test portions were taken from each. The test portions were labeled as "15A", "15B", "36A", "36B" etc. and quantitated via UV-VIS (Thermo Evolution 600) in the FTS laboratory facility. The samples were quantitated at 274.0 nm based on a four-point caffeine standard calibration curve that ranged from 27%-3.375%.

Stability was not tested as the samples are not reasonably expected to change when sealed in glass vials. Tests will be shipped in boxes with 'Refrigerate Upon Arrival' stickers to prevent damage and evaporation.

Statistical Analysis of Homogeneity Data:

Table 1: Item 1 Duplicated results for ten distribution units and intermediate stages of calculation in Cochran's test

| Item 1 Sample | A (%) | B (%) | D=A-B | S=A+B | D ² =(A-B) ² |
|---------------|--------|--------|--------|--------|------------------------------------|
| 15 | 26.851 | 27.013 | -0.162 | 53.864 | 0.0262 |
| 36 | 26.775 | 26.828 | -0.053 | 53.603 | 0.0028 |
| 45 | 26.983 | 26.783 | 0.200 | 53.766 | 0.0400 |
| 87 | 26.797 | 27.025 | -0.228 | 53.822 | 0.0520 |
| 106 | 26.766 | 26.771 | -0.005 | 53.537 | 0.0000 |
| 20 | 26.744 | 26.684 | 0.060 | 53.428 | 0.0036 |
| 42 | 26.735 | 26.900 | -0.165 | 53.635 | 0.0272 |
| 55 | 26.955 | 26.786 | 0.169 | 53.741 | 0.0286 |
| 81 | 26.703 | 26.702 | 0.001 | 53.405 | 0.0000 |
| 115 | 26.725 | 26.933 | -0.208 | 53.658 | 0.0433 |

Cochran's Test:

Analytical outliers should be deleted from the data before one-way analysis of variance (ANOVA) is carried out; Cochran's test is suitable.

Calculate the test statistic (C):

$$C = \frac{D_{MAX}^2}{\sum D_i^2}$$

$$= \frac{0.0520}{0.2237} = 0.232$$

Where C=Cochran's statistic test

D_{MAX}=the largest difference between duplicates

D_i=difference of each pair of duplicates

Table 2: Critical Values for the Cochran test statistic for duplicates

| m ¹ | 95% |
|----------------|-------|
| 10 | 0.602 |
| 11 | 0.570 |
| 12 | 0.541 |
| 13 | 0.515 |

Where m¹ = number of samples that have been measured in duplicate.

The 5% critical value for ten samples from Table 2 is 0.602.

In the data from Table 1, no analytical outliers were identified.

Estimate of Analytical and Sampling Variances:

One-way ANOVA is used to estimate the analytical and sampling variance and is performed in Excel.

Table 3: Item 1 One-way ANOVA Output

| Source of Variation | SS | df | MS | F | P-Value | F crit |
|---------------------|----------|----|----------|----------|----------|----------|
| Between Groups | 0.110762 | 9 | 0.012307 | 1.100244 | 0.438469 | 3.020383 |
| Within Groups | 0.111856 | 10 | 0.011186 | | | |

$$\text{So } S_{an}^2 = MS_{within} = 0.011186$$

where S_{an}^2 = the analytical variance

$$\text{and } S_{sam}^2 = \frac{MS_{between} - MS_{within}}{2}$$

$$= \frac{0.012307 - 0.011186}{2} = 0.00056$$

where S_{sam}^2 = the between sample variance

Test for Sufficient Analytical Precision:

The target standard deviation (σ_{pt}) is the product of the mean of all duplicate results (\bar{x}) and the between-laboratory coefficient of variation (CV) which is established by the study coordinator.

$$\begin{aligned}\sigma_{pt} &= \bar{x} * CV \\ &= 26.823 (0.07) \\ &= 1.88 \%\end{aligned}$$

The analytical standard deviation (S_{an}) is the square root of the analytical variance estimated from ANOVA above.

$$S_{an}/\sigma_{pt} = 0.106/1.88 = 0.056$$

This is less than the critical value of 0.5. The method is precise enough to detect significant inhomogeneity.

Test for Acceptable Between Sample Variance:

Calculate the allowable sampling variance (σ_{all}^2) as

$$\begin{aligned}\sigma_{all}^2 &= (0.3 \times \sigma_{pt})^2 \\ &= (0.3 \times 1.88)^2 \\ &= 0.318\end{aligned}$$

Where σ_{pt} = target standard deviation

The critical value is:

$$\begin{aligned}c &= F_1(\sigma_{all}^2) + F_2(S_{an}^2) \\ &= 1.88(0.318) + 1.01(0.011186) \\ &= 0.61\end{aligned}$$

The values for factors F_1 and F_2 are presented in Table 4.

Table 4: Factors F_1 and F_2 for use in testing for sufficient homogeneity

| m^1 | 13 | 12 | 11 | 10 |
|-------|------|------|------|------|
| F_1 | 1.75 | 1.79 | 1.83 | 1.88 |
| F_2 | 0.80 | 0.86 | 0.93 | 1.01 |

Where m^1 = number of samples that have been measured in duplicate.

Compare the sampling variance S_{sam}^2 with the critical value.

The sampling variance ($S_{sam}^2 = 0.00056$) is less than the critical value (0.61). The samples are sufficiently homogenous.

The results of the sufficient homogeneity testing are summarized in Table 5.

Table 5: Item 1 Homogeneity test results

| Item 1 | Value | Critical | Result |
|----------------------|---------|----------|--------|
| Cochran | 0.232 | 0.602 | Pass |
| S_{an}/σ_{pt} | 0.056 | 0.50 | Pass |
| S_{sam}^2 | 0.00056 | 0.61 | Pass |

Table 6: Item 2 Duplicated results for ten distribution units and intermediate stages of calculation in Cochran's test

| Item 2 Sample | A (%) | B (%) | D=A-B | S=A+B | $D^2=(A-B)^2$ |
|---------------|-------|-------|--------|-------|---------------|
| 23 | 4.868 | 4.870 | -0.002 | 9.738 | 0.0000 |
| 97 | 4.956 | 4.927 | 0.029 | 9.883 | 0.0008 |
| 120 | 4.869 | 4.882 | -0.013 | 9.751 | 0.0002 |
| 10 | 4.915 | 4.939 | -0.024 | 9.854 | 0.0006 |
| 82 | 4.864 | 4.909 | -0.045 | 9.773 | 0.0020 |
| 112 | 4.927 | 4.915 | 0.012 | 9.842 | 0.0001 |
| 86 | 4.857 | 4.868 | -0.011 | 9.725 | 0.0001 |
| 21 | 4.903 | 4.908 | -0.005 | 9.811 | 0.0000 |
| 92 | 4.864 | 4.874 | -0.010 | 9.738 | 0.0001 |
| 27 | 4.906 | 4.908 | -0.002 | 9.814 | 0.0000 |

Cochran's Test:

Analytical outliers should be deleted from the data before one-way analysis of variance (ANOVA) is carried out; Cochran's test is suitable.

Calculate the test statistic (C):

$$C = \frac{D_{MAX}^2}{\sum D_i^2}$$

$$C = \frac{0.0020}{0.039} = 0.05$$

Where C=Cochran's statistic test

D_{MAX} =the largest difference between duplicates

D_i =difference of each pair of duplicates

The 5% critical value for ten samples from Table 2 is 0.602.

In the data from Table 6, no analytical outliers were identified.

Estimate of Analytical and Sampling Variances:

One-way ANOVA is used to estimate the analytical and sampling variance and is performed in Excel.

Table 7: Item 2 One-way ANOVA Output

| Source of Variation | SS | df | MS | F | P-Value | F crit |
|---------------------|----------|----|----------|----------|---------|----------|
| Between Groups | 0.013912 | 9 | 0.001546 | 7.711787 | 0.00184 | 3.020383 |
| Within Groups | 0.002005 | 10 | 0.0002 | | | |

So $S_{an}^2 = MS_{within} = 0.0002$

where S_{an}^2 = the analytical variance

$$\text{and } S_{sam}^2 = \frac{MS_{between} - MS_{within}}{2}$$

$$= \frac{0.001546 - 0.0002}{2} = 0.00067$$

where S_{sam}^2 = the between sample variance

Test for Sufficient Analytical Precision:

The target standard deviation (σ_{pt}) is the product of the mean of all duplicate results (\bar{x}) and the between-laboratory coefficient of variation (CV) which is established by the study coordinator.

$$\begin{aligned} \sigma_{pt} &= \bar{x} (CV) \\ &= 4.897 (0.07) \\ &= 0.34 \% \end{aligned}$$

The analytical standard deviation (S_{an}) is the square root of the analytical variance estimated from ANOVA above.

$$S_{an}/\sigma_{pt} = 0.014/0.34 = 0.041$$

This is less than the critical value of 0.5. The method is precise enough to detect significant inhomogeneity.

Test for Acceptable Between Sample Variance:

Calculate the allowable sampling variance (σ_{all}^2) as

$$\begin{aligned}\sigma_{all}^2 &= (0.3 \times \sigma_{pt})^2 \\ &= (0.3 \times 0.34)^2 \\ &= 0.0104\end{aligned}$$

Where σ_{pt} = target standard deviation

The critical value is:

$$\begin{aligned}c &= F_1 (\sigma_{all}^2) + F_2 (S_{an}^2) \\ &= 1.88 (0.0104) + 1.01 (0.0002) \\ &= 0.0198\end{aligned}$$

The values for factors F_1 and F_2 are presented in Table 4.

Compare the sampling variance S_{sam}^2 with the critical value.

The sampling variance ($S_{sam}^2 = 0.00067$) is less than the critical value (0.0198). The samples are sufficiently homogenous.

The results of the sufficient homogeneity testing are summarized in Table 8.

Table 8: Item 2 Homogeneity test results

| Item 2 | Value | Critical | Result |
|----------------------|---------|----------|--------|
| Cochran | 0.05 | 0.602 | Pass |
| S_{an}/σ_{pt} | 0.041 | 0.50 | Pass |
| S_{sam}^2 | 0.00067 | 0.0198 | Pass |

Please quantitate the following items containing Ethanol.

Items Submitted

Item 1: Sample of liquid that contains Ethanol.

Item 2: Sample of liquid that contains Ethanol.

3) Indicate the Method and Parameters Used for Quantitation:

Example: GC-FID: column: SGE 12 m x 0.22 mm (F/N054046). 4-point calibration: 0 - 1.25 mg/mL.
Internal standard: Tetracosane, Solvent: Methanol.

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|--|
| p20252701 | W049 | HSS-GC/MS-FID, column: DB-ALC 123-9134, 30 m x 0.320 mm x 1.80 um, 4-point calibration, 0.020 to 0.500 g/100 mL. Internal standard: n-propanol, Solvent/diluent: deionized water. FID is used for ethanol retention time identification and quantitation, MS is used for ethanol spectral comparison. Calibration is linear regression with inverse weighing. Calibrator response factors before and after samples are averaged. |
| p20252702 | W163 | GC-FID Dual Column System. 6 Point Calibration: 10 - 400 mg%. ISTD n-propanol. |
| p20252703 | W088 | Instrument: HS-GC/MS-FID Column: Agilent DB-ALCI 0.32m 6 Point Calibrator Curve: 0.010-0.400 g/dL Internal Standard: n-propanol |
| p20252704 | W088 | Method parameters: Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6 point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252705 | W088 | 1. Instrument: HS-GC/MS-FID 2. Column: Agilent DB-ALC 1 0.32mm. 3. 6 point calibration: 0.010 - 0.400 g/dL 4. Internal Standard: n-propanol |
| p20252706 | W088 | Instrument: HS-GC-FID Column: Agilent DB5-ALCI 0.32mm 6-point calibration curve: 0.010 - 0.400 g/dL Internal Standard: n-propanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|--|
| p20252707 | W088 | The instrument used for testing was a Headspace GC-FID/MS. The column used was an Agilent 123-9134 with the dimension of 30 m x 320 μ m x 1.8 μ m. A 6-point calibration curve was utilized, which ranges from 0 - 0.40 g/100 mL. Lastly, the internal standard used was a 0.02 g% solution of n-propanol in DI water. |
| p20252708 | W088 | GC-FID Column: Agilent DB-Alc1 0.32 id 30m 6-point Calibration: 0.10-0.40 g/dL Internal Std: 0.02g/dL n-propanol |
| p20252709 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6 point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252710 | W088 | Instrument: HS-GC-MS Column: Agilent DB-ALC1 0.32 mm Calibration: 6-point calibration curve 0.010 - 0.400 g/dL Internal Standard: n-propanol |
| p20252711 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6 point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252712 | W088 | a. Instrument: HS-GC/MS-FID b. Column: Agilent DB-ALC1 0.32 mm c. 6 points calibrator curve: 0.010 - 0.400 g/dL d. Internal Standard: n-propanol |
| p20252713 | W088 | Instrument: HS GCMS FID Column: DB ALC1 0.32mm 6 point calibration 0.01-0.40g/dL internal standard n-propanol |
| p20252714 | W088 | HS-GC/MS-FID: column: Agilent DB-ALC1 0.32m. 6-point calibration curve 0.010-0.400 g/dL. Internal standard: n-propanol (0.02g/dL) |
| p20252715 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32mm 6 point calibration curve: 0.010-0.400 g/dL Internal Standard: n-propanol |
| p20252716 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32mm 6 point calibration curve: 0.010-0.400 g/dL Internal Standard: n-propanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|---|
| p20252717 | W088 | HS-GC-FID a. Instrument: HS-GC/MS-FID b. Column: Agilent DB-ALCI 0.32mm c. 6 point calibration curve: 0.010-0.400 g/dL d. Internal standard: n-propanol |
| p20252718 | W088 | HSGC-FID column: Agilent DB Alc1 0.32mm - 6 point calibration: 0.010 - 0.400 g/100mL internal standard: 1-propanol |
| p20252719 | W088 | instrument: HS-GC-FID Column: DB-ALCI 0.32 mm 6 point calibration curve: 0.010 - 0.400 g/dL internal standard: n-propanol |
| p20252720 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32mm 6 point calibration curve: 0.010-0.400 g/dL Internal standard: n-propanol |
| p20252721 | W088 | HS-GC-FID-MS. Agilent DB-ALC1 0.32mm column, 6 point calibration: 0.010-0.400 g/100mL. internal standard: n-propanol |
| p20252722 | W088 | HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL internal standard: n-propanol |
| p20252723 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252724 | W088 | Instrument: HS-GC-FID/MS Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252725 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|--|
| p20252726 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252727 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252728 | W088 | Instrument: HS-GC-FID Column: Agilent DB-ALCI 0.32 mm 6-point calibration curve: 0.010-0.400 g/dL Internal std: n-propanol |
| p20252733 | W092 | GC-FID: dual columns: DB-BAC 1 (30m, 0.32mm ID FT, 1.2 μ m) and DB-BAC 2 (30m, 0.32mm ID FT, 1.8 μ m) respectively coupled with two independent flame ionization detectors. 5-point calibration: 1-50% v/v. Internal standard: n-Propyl alcohol Solvent: D.I. water |
| p20252735 | W031 | GC-FID Column A: Restek Rtx-BAC PLUS 1 (30 m, 0.32 mm ID, 1.80 μ m) (Serial # 110094, Cat # 18006) Column B: Restek Rtx-BAC PLUS 2 (30 m, 0.32 mm ID, 0.6 μ m) (Serial #1104092, Cat # 18004) 3-point calibration: 0 - 400 mg/dL Internal Standard: N-propanol (Sigma-Aldrich Lot# SHBM7906) Solvent: deionized water |
| p20252736 | W079 | GC-FID Glass packed column: 80/120 Carbopack B/5% CARBOWAX 20M (1.8m x 6.35mm) Semi-Quant Solvent: Deionized water |
| p20252737 | W079 | GC-FID Glass packed column: 80/120 Carbopack B/5% CARBOWAX 20M (1.8m x 6.35mm) Semi-Quant Solvent: Deionized water |
| p20252738 | W044 | GC-FID Column 1: 30 m Rtx-BAC Plus 1 (30 m x 320 μ m x 1.8 μ m) Column 2: 30 m Rtx-BAC Plus 2 (30 m x 320 μ m x 0.6 μ m) 7-point calibration: 0.01-0.500 g/100 mL Internal Standard: N-Propanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|---|
| p20252739 | W114 | a. GC-FID b. Columns: DB-ALC1 and DB-ALC2 c. 5-point calibration: 0.01-0.4 g/dL d. Internal standard: n-Propanol |
| p20252740 | W114 | GC-FID: columns: DB-ALC1 and DB-ALC2. 5-point calibration: 0.01-0.4 g/dL. Internal Standard: n-Propanol |
| p20252741 | W114 | GC-FID: columns: DB-ALC1 and DB-ALC2. 5-point calibration: 0.01-0.4 g/dL. Internal standard: n-Propanol. |
| p20252742 | W114 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point calibration: 0.01 - 0.4 g/dL Internal standard: n-Propanol |
| p20252743 | W114 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point clibration: 0.01-0.4 g/dL Internal standard: n-propanol |
| p20252744 | W114 | GC-FID 5-point calibration: 0.01 - 0.4 g/dL Columns: DB-ALC1 and DB-ALC2 Internal standard: n-Propanol |
| p20252745 | W114 | a. GC-FID b. Columns: DB-ALC1 and DB-ALC2 c. 5-point calibration: 0.01-0.4 g/dL d. Internal Standard: n-Propanol |
| p20252746 | W114 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point calibration: 0.01 - 0.4 g/dL Internal standard: n-Propanol |
| p20252747 | W114 | GC-FID, columns DB-ALC1 and DB-ALC2, five point calibration curve:0.01 g/dL to 0.40 g/dL. Internal standard: n-propanol |
| p20252749 | W190 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point calibration: 0.01 - 0.4 g/dL Internal standard: n-Propanol |
| p20252750 | W190 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point calibration: 0.01-0.4 g/dL Internal standard: n-Propanol |
| p20252751 | W190 | GC-FID Columns: DB-ALC1 and DB-ALC2 5-point calibration: 0.01 - 0.4 g/dL Internal standard: n-Propanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|--|
| p20252752 | W190 | a. GC-FID b. Columns: DB-ALC1 and DB-ALC2 c. 5-point calibration: 0.01-0.4 g/dL d. Internal standard: n-Propanol |
| p20252753 | W190 | 5-point calibration: 0.01 - 0.4 g/dL GC-FID Columns: DB-ALC1 and DB-ALC2 Internal standard: n-propanol |
| p20252754 | W190 | GC-FID: column: DB-ALC1 and DB-ALC2, 5-point calibration:0.01-0.4 g/dL. Internal standard: N-propanol |
| p20252755 | W190 | GC-FID Columns: DB-ALC1 and DB-ALC2 5 point calibration 0.01-0.4 g/dL Internal standard: n-propanol |
| p20252756 | W190 | Instrument: GC-FID Columns: DB-ALC1 & DB-ALC2 5-point calibration: 0.01-0.4 g/dL Internal Standard: n-Propanol |
| p20252757 | W190 | a. GC-FID b. Columns: DB-ALC1 and DB-ALC2 c. 5-point calibration: 0.01-0.4 g/dL d. Internal standard: n-Propanol |
| p20252758 | W002 | 7697A Headspace Agilent auto-sampler to Intuvo 900 GC System with dual columns DB-BAC1 and DB-BAC2 (UI: 30m, 0.32mm, 1.2um). 6-point calibration from 20 mg/dL to 500 mg/dL. |
| p20252761 | W110 | HS-GC-FID: 30m x 0.32 mm ID, 1.8 df, Rtx-BAC1 capillary column. 4-point calibration: 0.040, 0.100, 0.200, 0.400 g/100mL. Internal Standard: n-Propanol. |
| p20252762 | W110 | HS-GC-FID: 30 m x 0.32 mm ID, 1.8 df, Rtx-BAC1 capillary column. 4-point calibration: 0.040, 0.100, 0.200 and 0.400 %. Internal standard: n-Propanol. |
| p20252766 | W110 | H-GC-FID: 30m x 0.32mm ID, 1.8 df, Rtx-BAX1 capillary column, 4-point calibration: 0.040, 0.100, 0.200, and 0.400%. Internal standard n-Propanol. |
| p20252769 | W110 | HS-GC-FID: L 30m x 0.32mm ID, 1.8 df, Rtx-BAC1 capillary column, 4-point calibration: 0.040%, 0.100%, 0.200%, 0.400%. Internal Standard n-propanol. |
| p20252770 | W082 | GC-FID: column: BAC+1 / BAC+2, 4-point calibration (linear, not forced to zero): 0.010 - 0.400 g/dL. Internal standard: n-propanol |
| p20252771 | W172 | Head space GC FID and Head space GC MS/FID |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|---|
| p20252772 | W096 | GC-FID 30 m x 0.32 mm x 1.8 μ m BAC-1 Plus Restek Column (quantitative and qualitative analysis) GC-FID 30 m x 0.32 mm x 0.6 μ m BAC-2 Plus Restek Column (qualitative analysis only) 5-point calibration curve from 0.50 g/dL to 20 g/dL; 4-point calibration curve from 20 g/dL to 80 g/dL. Internal Standard - 0.4% n-propanol in water |
| p20252775 | W206 | GC-FID: Column: Restek BACPlus-2. 0.020-0.500 g%. Internal standard: T-Butanol |
| p20252776 | W116 | GC-FID-MS: FID Column: DB-ALC1 30 m X 0.32 mm X 1.80 μ m (Agilent 123-9134), MS Column: DB-1701 30 m X 0.25 mm X 1 μ m (Agilent 122-0733). 6-point calibration:0.010 - 0.500 g/100mL. Internal standard: n-propanol, Solvent: DI water. |
| p20252777 | W116 | GC-FID-MS: FID column: DB-ALC1 30m x 0.32mm x 1.8 μ m (Agilent 123-9134), MS column: DB-1701 30m x 0.25mm x 1 μ m (Agilent 122-0733), 6-point calibration: 0.010-0.500 g/100mL. Internal standard: n-Propanol, Solvent: Water |
| p20252778 | W025 | Mettler Toledo Density Meter DM50, 2-point calibration, Internal Standard: Air/Water 0.00119/0.99820 Adjustment Factor 1.516983 |
| p20252779 | W256 | GC-FID: column: Restek Rtx-BAC1 30 m x 0.53 mm (P/N - 18001). 7-point calibration: 10-400 mg/dL. Internal standard: n-propanol, Solvent: water with ammonium sulfate. |
| p20252780 | W070 | GC-MS column Restek BAC2 , GC-FID column Restek BAC1, 7 point calibration curve 0.01 - 0.5 gm%, Internal Standard 0.01667 v/v% n-Propanol, Solvent: DI Water |
| p20252781 | W070 | GC-MS column Restek 18002, GC-FID column Restek 18003, 7 point calibration 0.01-0.50g%, Internal Standard 0.01667% N-Propanol, Solvent: Water |
| p20252782 | W070 | GC-MS column Restek 18002 (Rtx-BAC2); GC-FID column Restek 18003 (Rtx-BAC1), 5 point calibration 0.01-0.50g%, Internal Standard 0.01667% N-Propanol, Solvent: Water |
| p20252783 | W070 | GC-MS column Restek 18002, GC-FID column Restek 18003, 5-point calibration: 0.01-0.50 gm%, Internal standard: 0.01667% N-Propanol, Solvent: Water. |
| p20252784 | W055 | GC-FID Column: Restek Rxi-1ms 30m x 0.25mm 4-point calibration: 0.32-4.06 mg/ml Internal standard: n-Butanol Reference material: 95% Ethanol by volume |
| p20252785 | W055 | GC-FID Column RX-1MS - 0.25mm x 0.25 μ m x 30m 4 point calibration 0.32mg/mL, 1.62mg/mL, 2.44mg/mL, 4.06mg/mL Internal standard: butanol |

| UTIC | Webcode | Indicate the Method and Parameters Used for Quantitation |
|-----------|---------|---|
| p20252786 | W055 | GC-FID, 4-point calibration: 0.32-4.06 mg/mL. Internal standard: n-butanol, Solvent: DI Water |
| p20252787 | W055 | GC-FID Column: Restek Rxi-1ms 30m x 0.25mm 4-point calibration: 0.32-4.06 mg/ml Internal standard: n-Butanol Reference material: 95% Ethanol by volume |
| p20252788 | W004 | Headspace GC-FID: column: DB-ALC1 (30 m x 0.320 mm x 1.8 μ m). 4-point calibration: 0.5%-50% v/v Internal Standard: 0.5% (v/v) n-propanol. Solvent: deionized water |
| p20252789 | W037 | HS-GC-FID-MS: column: Agilent DB-ALC1 30mx0.320mm i.d. x 1.80um film. 5-point calibration: 0.01 g/100ml -0.40 g/100ml. Internal Standard: N-Propanol ISTD Solution. |

4) What standard was used and, if given, it's stated purity?

Example: Sigma-Aldrich Batch 079K1053 96.5%

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|--|
| p20252701 | W049 | Calibrator/Control Information Level Provider Lot Expiration 0.020 g/100 mL Supelco FN03122113 Exp 03/2026 0.100 g/100 mL Supelco FN04042408 Exp 04/2029 0.200 g/100 mL Supelco FN03132302 Exp 03/2028 0.500 g/100 mL Supelco FN01172402 Exp 01/2029 Mixed Volatiles Cerilliant FN02162310 Exp 05/2028 0.025 g/100 mL Lipomed 30102019-C Exp 05/2026 Beverage 14% ABV SCDL 3 Exp None 0.300 g/100 mL Lipomed 530442 Exp 08/2027 |
| p20252702 | W163 | Certified Supelco standards for calibrators. Certified LGC standards for QCs. |
| p20252703 | W088 | Calibration Kit: Supelco Multicomponent Alcohol Calibration Kit C1-C6 (MeOH, EtOH, Acetone, 2-Propanol), FN05092202, >99.9% Control: Supelco 0.05 g% Mix Control (MeOH, EtOH, Acetone, 2-Propanol), FN01212104, >99.9% |
| p20252704 | W088 | Calibrators: Cerilliant Lot FN05092202 >99.9% Validators: Cerilliant Lot FN09092101 >99.9%, Cerilliant Lot FN07252344 >99.9%, Cerilliant Lot FN06232204 >99.9% Controls: Cerilliant Lot FN05232301 >99.9%, Cerilliant Lot FN02162310 >99.9%, Cerilliant Lot FN06112149 >99.9% |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252705 | W088 | <p>Calibration Kit: Ceriliant Mix Calibration Kit (MeOH, EtOH, Acetone, 2-Propanol) : FN05092202; >99.9% (0.01, 0.025, 0.05, 0.10, 0.20, 0.40g%)</p> <p>Controls:</p> <ol style="list-style-type: none"> 1. Ceriliant 0.05 g% Mix Control (MeOH, EtOH, Acetone, 2-Propanol): FN05232301; >99.9% 2. Ceriliant 0.10 g% Mix Control (MeOH, EtOH, Acetone, 2-Propanol): FN02162310; >99.9% 3. Ceriliant 0.30 g% Ethanol Control (EtOH): FN06112149 ; >99.9% |
| p20252706 | W088 | <p>Ceriliant Multicomponent Alcohol Calibration Kit (01-06) - NIST Traceable, Solution Lot: FN05092202</p> <p>C1 FN01042201 Level 100 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> <p>C2 FN01042202 Level 250 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> <p>C3 FN01042203 Level 500 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> <p>C4 FN01042204 Level 1000 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> <p>C5 FN01042205 Level 2000 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> <p>C6 FN05062202 Level 4000 mcg/mL: purity ≥ 99.9% for Acetone, Methanol, Ethanol, and Isopropanol</p> |
| p20252707 | W088 | <p>Three standards were used for testing. The 0.10 g% Multicomponent Alcohol Mix (Ceriliant Solution Lot FN02162310) had a purity of 99.9% for all components. The 0.05 g% Multicomponent Alcohol Mix (Ceriliant Solution Lot FN05232301) had a purity of 99.9% for all components. The 0.30 g% Ethanol Solution (Ceriliant Solution Lot FN06112149) had a purity of 99.9%.</p> |
| p20252708 | W088 | <p>Cal: Supelco FN05092202 ≥99.9% Val: Supelco FN07252344 >99.9%</p> <p>Controls: Supelco FN05232301 >99.9%; FN02161210 >99.9%, FN06112149 >99.9%</p> |
| p20252709 | W088 | <p>Ceriliant for all analytes, Calibration: Multicomponent Alcohol Calibration Kit (C1 - C6) FN05092202 >99.9% Validation: Multicomponent Alcohol Calibration Kit FN07252344 >99.9% Multicomponent Alcohol Mix 100 FN04162405 >99.9% Ethanol-80 FN0623204 >99.9% Controls: Multicomponent Alcohol Mix 500 FN05232301 >99.9% Multicomponent Alcohol Mix 1000 FN02162310 >99.9% Ethanol-300 FN06112149 >99.9%</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|--|
| p20252710 | W088 | <p>Calibration (Supelco)</p> <p>- Calibration Kit (Methyl Alcohol, Ethyl Alcohol, Isopropyl Alcohol, Acetone); Lot # FN05092202; Expiration Date: 05/2027; Purity: ≥99.9%</p> <p>Controls (Supelco)</p> <p>- 0.10 g/dL mix (Methyl Alcohol, Ethyl Alcohol, Isopropyl Alcohol, Acetone); Lot # FN02162310; Expiration Date: 05/2028; Purity: >99.9%</p> <p>- 0.30 g/dL Ethyl Alcohol; Lot # FN06112149; Expiration Date: 06/2026; Purity: >99.9%</p> <p>- 0.05 g/dL mix (Methyl Alcohol, Ethyl Alcohol, Isopropyl Alcohol, Acetone); Lot # FN05232301; Expiration Date: 05/2028; Purity: >99.9%</p> |
| p20252711 | W088 | <p>Calibration: Supelco FN05092202 exp. 05/2027 Methyl Alcohol, Ethyl Alcohol, Isopropyl Alcohol, Acetone ≥99.9%</p> <p>Verification: Supelco FN07252344 exp. 04/2028 Methyl Alcohol, Isopropyl Alcohol, Acetone >99.9% Ethyl Alcohol 99.9%</p> <p>Supelco FN04162405 exp. 04/2029 Methyl Alcohol, Ethyl Alcohol, Isopropyl Alcohol, Acetone >99.9%</p> <p>Supelco FN06232204 exp. 04/2027 Ethyl Alcohol >99.9%</p> <p>Controls: Supelco FN05232301 exp.05/2028 Methyl Alcohol, Isopropyl Alcohol, Acetone >99.9% Ethyl Alcohol 99.9%</p> <p>Supelco FN02162310 exp 05/2028 Methyl Alcohol, Isopropyl Alcohol, Acetone >99.9% Ethyl Alcohol 99.9%</p> <p>Supelco FN06112149 exp 06/2028 Ethyl Alcohol >99.9%</p> |
| p20252712 | W088 | <p>Calibration kit: Cerilliant Mix Calibration Kit (MeOH, EtOH, Acetone, 2-Propanol): FN08182003; > 99.9% (0.010 g/%) and FN05092202; 99.9% (0.025, 0.050, 0.100, 0.200, 0.400 g/%)</p> <p>Controls:</p> <ul style="list-style-type: none"> • Cerilliant 0.05 g% Mix Control (MeOH, EtOH, Acetone, 2-Propanol): FN05232301; >99.9% • Cerilliant 0.10 g% Mix Control (MetOH, EtOH, Acetone, 2-Propanol): FN02162310; >99.9% • Cerilliant 0.30 g% Ethanol Control (EtOH): FN06112149; >99.9% |
| p20252713 | W088 | <p>Supelco for all standards calibration kit (EtOH, MeOH, IPA, Ace) FN05092202 0.01, 0.025, 0.05, 0.10, 0.20, 0.40 g/dL >99%</p> <p>Controls for run</p> <p>.05 g/dL Mix (MeOH, EtOH, IPA, Ace) FN01212104 >99%</p> <p>0.10 g/dL Mix (MeOH, EtOH, IPA, Ace) FN02162310 >99%</p> <p>0.30 g/dL EtOH FN06112149 >99%</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|--|
| p20252714 | W088 | <p>Calibration: Cerilliant Calibration Mix Kit: (MeOH, EtOH, Acetone, 2-propanol): FN05092202 >99.9%</p> <p>Controls: Cerilliant 0.05 g/dL Mix Control (FN05232301); Cerilliant 0.10 g/dL Mix (FN02162310); Cerilliant 0.30 g/dL EtOH (FN06112149). All >99.9%</p> |
| p20252715 | W088 | <p>All stated 99.9% (or greater) purity</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> |
| p20252716 | W088 | <p>All stated 99.9% (or greater) purity</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|--|
| p20252717 | W088 | <p>All stated 99.9% (or greater) purity</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> |
| p20252718 | W088 | <p>Cerilliant A-127 kit FN05092202 Cerilliant A-076 FN04162405 Cerilliant A-054 kit FN07252344 Cerilliant E-030 FN06232204 Cerilliant E-033 FN06112149 Cerilliant A-056 FN02162310 Cerilliant A-057 FN05232301 all 99.9%</p> |
| p20252719 | W088 | <p>All stated 99.9% (or greater) purity</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310)</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252720 | W088 | <p>All stated 99.9% (or greater) purity</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN03062415) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> |
| p20252721 | W088 | <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> <p>Calibration Cerilliant A-127 kit (FN05092202)</p> <p>Validation Cerilliant A-076 (FN04162405) Cerilliant A-054 kit (FN07252344) Cerilliant E-030 (FN06232204)</p> <p>Controls Cerilliant E-033 (FN06112149) Cerilliant A-056 (FN02162310) Cerilliant A-057 (FN05232301)</p> <p>All >99.9% purity</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? | | | | | | | | | | | | | | | | | | | | | |
|-------------|------------|---|-------------|------------|----------|-----------|------------|--|-----------|------------|------------|-----------|------------|------------|-----------|------------|------------|-----------|--|--|------------|------------|--|
| p20252722 | W088 | <p>Cerilliant for all analytes, all greater than 99.9% pure</p> <table border="1"> <thead> <tr> <th>Calibrators</th><th>Validators</th><th>Controls</th></tr> </thead> <tbody> <tr> <td>FN1042201</td><td>FN09082101</td><td></td></tr> <tr> <td>FN1042202</td><td>FN04032301</td><td>FN02162310</td></tr> <tr> <td>FN1042203</td><td>FN04032302</td><td>FN03062415</td></tr> <tr> <td>FN1042204</td><td>FN04032303</td><td>FN05232301</td></tr> <tr> <td>FN1042205</td><td></td><td></td></tr> <tr> <td>FN05062202</td><td>FN06232204</td><td></td></tr> </tbody> </table> | Calibrators | Validators | Controls | FN1042201 | FN09082101 | | FN1042202 | FN04032301 | FN02162310 | FN1042203 | FN04032302 | FN03062415 | FN1042204 | FN04032303 | FN05232301 | FN1042205 | | | FN05062202 | FN06232204 | |
| Calibrators | Validators | Controls | | | | | | | | | | | | | | | | | | | | | |
| FN1042201 | FN09082101 | | | | | | | | | | | | | | | | | | | | | | |
| FN1042202 | FN04032301 | FN02162310 | | | | | | | | | | | | | | | | | | | | | |
| FN1042203 | FN04032302 | FN03062415 | | | | | | | | | | | | | | | | | | | | | |
| FN1042204 | FN04032303 | FN05232301 | | | | | | | | | | | | | | | | | | | | | |
| FN1042205 | | | | | | | | | | | | | | | | | | | | | | | |
| FN05062202 | FN06232204 | | | | | | | | | | | | | | | | | | | | | | |
| p20252723 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> | | | | | | | | | | | | | | | | | | | | | |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252724 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> |
| p20252725 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252726 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> |
| p20252727 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252728 | W088 | <p>Calibrators: Cerilliant for all analytes all >99.9% pure</p> <p>FN01042201 FN01042202 FN01042203 FN01042204 FN01042205 FN05062202</p> <p>Validators: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN09082101 FN04032301 FN04032302 FN04032303 FN06232204</p> <p>Controls: Cerilliant for all analytes, all greater than 99.9% pure</p> <p>FN02162310 FN03062415 FN05232301</p> |
| p20252733 | W092 | <p>VWR Chemicals Ethanol absolute (99.7%)</p> <p>LGC Reference Spirit-5% ABV (Certified Reference Material)</p> |
| p20252735 | W031 | <p>The following NIST Traceable Certified Reference Standards were used:</p> <p>200 ± 1 mg/dL Ethanol, Cerilliant Lot FN03132302, >99.9%</p> <p>300 ± 2 mg/dL Ethanol, Cerilliant Lot FN06112149, >99.9%</p> |
| p20252736 | W079 | <p>Lipomed Ethanol 400.0 mg/dL</p> <p>Lot #12052023-B</p> |
| p20252737 | W079 | <p>Lipomed Lot No. 12052023-B, 400 mg/dL</p> |
| p20252738 | W044 | <p>Lipomed, Aqueous Ethanol Standard Solution 400 mg/dL, Lot # 12052023-B, 100.0% purity</p> |
| p20252739 | W114 | <p>Supelco Multicomponent Alcohol Calibration Kit lot# FN05092202 (purity > 99.9%) was used to calibrate the GC-FID.</p> <p>Alcohol Pharmaceutical Secondary Standard Lot# LRAD5063 (purity 92.7%) was used to make the in-house 0.080 g/dL whole blood control and the in-house 0.300 g/dL high aqueous control.</p> |
| p20252740 | W114 | <p>Cerilliant Calibration Kit Lot# FN05092202</p> |
| p20252741 | W114 | <p>Cerilliant Multicomponent Alcohol Calibration Kit (C1-C6) A-127-6X1.2ML Lot FN05092202</p> |
| p20252742 | W114 | <p>Cerilliant calibration kit Lot#: FN05092202</p> |
| p20252743 | W114 | <p>Cerilliant Multicomponent Alcohol Calibration Kit A-127 Lot FN05092202</p> |
| p20252744 | W114 | <p>Cerilliant Multicomponent Calibration Kit Lot# FN05092202</p> <p>Chromatographic purity of each neat component is >=99.9%</p> |
| p20252745 | W114 | <p>Cerilliant Calibration Kit: FN05092202</p> |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|--|
| p20252746 | W114 | Cerilliant Multicomponent Calibration Kit Lot# FN05092202 Chromatographic purity of each neat component is >=99.9% |
| p20252747 | W114 | Cerilliant multi component alcohol calibration kit, catalog A-127, lot FN05092202 |
| p20252749 | W190 | Cerilliant calibration kit: Lot# FN05092202, exp 5-31-27 |
| p20252750 | W190 | Cerilliant Certified Reference Material: 80 mg/dL and 300 mg/dL Cerilliant calibration kit: FN05092202 |
| p20252751 | W190 | Cerilliant Calibration Kit FN05092202 >99.9% |
| p20252752 | W190 | Cerilliant calibration kit FN05092202 |
| p20252753 | W190 | Cerilliant calibration kit Lot # FN05092202 exp. date 05.31.2027 |
| p20252754 | W190 | Cerilliant multi component alcohol calibration kit FN05092202 |
| p20252755 | W190 | Cerilliant calibration kit LOT FN05092202 |
| p20252756 | W190 | Cerilliant Multi-Component Alcohol Calibration Kit: FN05092202 |
| p20252757 | W190 | Multicomponent Alcohol Calibration Kit Lot# FN05092202, Cerilliant calibration kit |
| p20252758 | W002 | Supelco ethanol standards with purity 99.9% |
| p20252761 | W110 | Cerilliant Ethanol-80 Lot# FN06232204. Purity > 99.9% |
| p20252762 | W110 | Cerilliant Ethanol-80 Lot #FN06232204. Purity > 99.9% |
| p20252766 | W110 | Cerilliant Ethanol-80 lot #FN06232204, 99.9% |
| p20252769 | W110 | Cerilliant Ethanol-80 Lot# FN06232204 Purity > 99.9% |
| p20252770 | W082 | <u>Calibrators</u> Cerilliant FN12272201 0.010 g% Cerilliant FN03072301 0.100 g% Cerilliant FN03132302 0.200 g% Cerilliant FN03052102 0.400 g% |
| | | <u>Controls</u> Cerilliant FN03122113 0.020 g% Cerilliant FN06232204 0.080 g% Cerilliant FN06112149 0.300 g% |
| p20252771 | W172 | Cerilliant |
| p20252772 | W096 | Internal standard - n-propanol (Millipore-Sigma lot SHBN6629) Calibrators - ethanol (Pharmco-Aaper lot 240208-B079607) Controls - ethanol (Pharmco-Aaper lot 240208-B079608) |
| p20252775 | W206 | Ethanol 200 proof - Lot# SHBC2261V |
| p20252776 | W116 | Cerilliant 0.080 g/100mL Lot #:FN06232204 purity:>99.9% |
| p20252777 | W116 | Cerilliant 0.080 g/100mL Lot #:FN06232204 purity: >99.9% |
| p20252778 | W025 | Ethanol (400 +/- 2mg/dL) Sigma Aldrich (Supelco) lot FN03052102, >99.9% |
| | | Cerilliant Multicomponent alcohol calibration kit, Lot: FN07252344. 99.9% |

| UTIC | Webcode | What standard was used and, if given, it's stated purity? |
|-----------|---------|---|
| p20252780 | W070 | 0.01 g% MIAE Lipomed lot: 1808.1B1.1L11 (MIAE = Methanol, Isopropanol, Acetone, Ethanol) 0.03 g% MIAE Lipomed lot: 524303 0.04 g% MIAE Lipomed lot: 524308 0.05 g% MIAE Lipomed lot: 1808.1B1.1L10 0.10 g% MIAE Lipomed lot: 1808.1B1.1L13 0.20 g% MIAE Lipomed lot: 529866 0.50 g% MIAE Lipomed lot: 531813 NIST 0.08 g% Ceriliant lot: FN06232204 0.10 g% control Ceriliant lot: FN02162310 |
| p20252781 | W070 | Lipomed used for calibrators. 0.01 g% Lot 1808.1B1.1L11 , 0.03 g% Lot 524303, 0.04 g% Lot 524308, 0.05 g% Lot 1808.1B1.1L10, 0.10 g% Lot: 1808.1B1.1L13, 0.2 g% Lot 529866. Calibrator at 0.5 g% made in house using Thermo fisher Ethanol B0547734A (retest 03/31/29) 99.5% Ceriliant used for controls. 0.08 g% Lot FN06232204 and 0.1 g% Lot FN02162310 |
| p20252782 | W070 | Thermo Scientific Lot #: B0547734A, 99.5% |
| p20252783 | W070 | Lipomed- Alcohol mixture 4, Lot #1808.1B1.1L11, Lot #,524303, Lot #524308, Lot #534932, Lot # 1808.1B1.1L13, Lot # 529866, Lot # 531813. Ceriliant multicomponent mix (1000 ug/mL) Lot # FN02162310 , Ceriliant NIST Ethanol-80 Lot # FN06232204 |
| p20252784 | W055 | 95% Ethanol: Sigma-Aldrich, 190 Proof, 95%, Lot#SHBR8294 1-Butanol: Sigma-Aldrich, 99.9%, Lot#SHBS0450 |
| p20252785 | W055 | Ethanol 95% Sigma-Aldrich Lot: SHBR8294 |
| p20252786 | W055 | Sigma SHBR8294 95% |
| p20252787 | W055 | Ethanol 95% Sigma-Aldrich Lot: SHBR8294 |
| p20252788 | W004 | NIST Ethanol-water solution, Nominal Mass Fraction 95.6% (Standard Reference Material (SRM) 2900) |
| p20252789 | W037 | Calibration Standards: Ceriliant Ethanol 10mg/dl Lot FN12272201 Ceriliant Ethanol 20mg/dl Lot FN03122113 Ceriliant Ethanol 50mg/dl Lot FN03172301 Ceriliant Ethanol 100mg/dl Lot FN03072301 Ceriliant Ethanol 300mg/dl Lot FN06112149 Ceriliant Ethanol 400mg/dl Lot FN04092401 Lipomed Standards for positive control: Lipomed 80mg/dl Lot 517757 Lipomed 150mg/dl Lot 518346 |

5) Method for Estimating Uncertainty:

Example: Uncertainty estimate is based on method reproducibility, CRM uncertainty and sample homogeneity.

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|--|
| p20252701 | W049 | Uncertainty estimate is based on calibrator and control manufacturing variability, matrix-matched control reproducibility, and allowed sample reproducibility limits. |
| p20252702 | W163 | Measurement uncertainty measurement is based on method reproducibility and the uncertainty of equipment for (dilutor, pipettes, etc) and CRM |
| p20252703 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252704 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252705 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252706 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibration. |
| p20252707 | W088 | The estimations for measurement uncertainty are based on the precision and reproducibility of the testing method for all analytes. Each step of the testing process is taken into account including, but not limited to, pipetting, human factors, and instrumentation when considering the reproducibility. |
| p20252708 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252709 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252710 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252711 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252712 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|---|
| p20252713 | W088 | Uncertainty is based on controls charts of method and process reproducibility, as well as the purity of standards used for calibration solutions. |
| p20252714 | W088 | Measurement uncertainty is based on control charts of the method reproducibility, and purity of standards, pipettes, and volumetric glassware that were used for testing. |
| p20252715 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions |
| p20252716 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions |
| p20252717 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252718 | W088 | Uncertainty estimate is based on control charts of the method, process reproducibility and purity of the standards used to make calibration solutions. |
| p20252719 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility and the purity of standards used to make calibrator solutions. |
| p20252720 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252721 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252722 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252723 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252724 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252725 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252726 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|--|
| p20252727 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252728 | W088 | Uncertainty estimate is based on control charts of method and process reproducibility, and the purity of standards used to make calibrator solutions. |
| p20252733 | W092 | Uncertainty estimate is based on recovery, precision, purity of standard and standard preparation. |
| p20252735 | W031 | Uncertainty estimate is based on uncertainties of the Hamilton Diluter and Eppendorf pipette calibrations, CRM uncertainties, and method reproducibility. The following NIST Traceable Certified Reference Standard is used as a method reproducibility control: $150 \pm 1 \text{ mg/dL}$ Ethanol, Cerilliant Lot FN 07022010, >99.9% |
| p20252736 | W079 | Uncertainty estimate is based on the overall procedure and analysis |
| p20252737 | W079 | Uncertainty estimate is based on the overall procedure and analysts |
| p20252738 | W044 | Uncertainty of Measurement is based on measurement reproducibility, CRM uncertainty, duplicate aliquots, and the use of pipettes. |
| p20252739 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252740 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252741 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, diluter, pipettes, etc. |
| p20252742 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252743 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252744 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment (certified reference material, dilutor, pipettes, etc.) |
| p20252745 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252746 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|--|
| p20252747 | W114 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252749 | W190 | Measurment Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252750 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252751 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252752 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252753 | W190 | Mesurment Uncertainty estimate is based on the following components: method reproducibility, dublicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252754 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equiptment: CRM, dilutor, pipettes, etc. |
| p20252755 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, diluter, pipettes, flasks |
| p20252756 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipement: CRM, dilutor, pipettes, etc. |
| p20252757 | W190 | Measurement Uncertainty estimate is based on the following budget components: method reproducibility, duplicate analysis, and equipment: CRM, dilutor, pipettes, etc. |
| p20252758 | W002 | Uncertainty is based on method reproducibility, CRM of calibrators, and pipette calibration. |
| p20252761 | W110 | All components of uncertainty (i.e. standards and equipment), when analyzing forensic alcohol samples, are used to make a reasonable estimation of the uncertainty of measurement. |
| p20252762 | W110 | All components of uncertainty (i.e. standards and equipment), when analyzing forensic alcohol samples, are used to make a reasonable estimation of the uncertainty of measurement. |
| p20252766 | W110 | All components of uncertainty (i.e. standards and equipment), when analyzing forensic alcohol samples, are used to make a reasonable estimation of the uncertainty of measurement. |

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|--|
| p20252769 | W110 | All components of uncertainty (i.e. standards and equipment), when analyzing forensic alcohol samples, are used to make a reasonable estimation of the uncertainty of measurement. |
| p20252770 | W082 | Estimation of uncertainty for this method is based upon both a Type A and Type B evaluation. The Type A evaluation is covered by the reproducibility of the method and measurements of the positive control standards. The Type B evaluation included the equipment used in the method, the laboratory requirement of 5% agreement between replicates, and the purity of the reference material. |
| p20252771 | W172 | Uncertainty estimate is based on method reproducibility. CRM uncertainty and sample homogeneity. |
| p20252772 | W096 | Uncertainty estimate based on method reproducibility, diluter/dispenser uncertainty, and glassware uncertainty. |
| p20252775 | W206 | Uncertainty estimate is based on duplicate sampling criteria and historical QC data. CRM, glassware and pipette uncertainty is also considered. Expanded uncertainty = 8% at 95.45% confidence interval (k=2) |
| p20252776 | W116 | Uncertainty estimate is based on method reproducibility. The following are included in the calculation: measurement reproducibility (control chart), CRM uncertainty for calibrators, internal standard dispensette uncertainty, matrix (duplicate result acceptability), dilution (volumetric flask uncertainty), pipette uncertainty. |
| p20252777 | W116 | Uncertainty estimate is based on method reproducibility. The following are included in the calculation: measurement reproducibility (control chart), CRM uncertainty for calibrators, internal standard dispensette uncertainty, matrix (duplicate result acceptability), dilution (volumetric flask uncertainty), pipette uncertainty. |
| p20252778 | W025 | Uncertainty of the densitometer, reference standard uncertainty, uncertainty of the process, instrument bias |
| p20252779 | W256 | Uncertainty is based on known QC recovery, CRM uncertainty and dispenser/pipette coefficient of variation. |
| p20252780 | W070 | Uncertainty estimate is based on method reproducibility based off of historic control data, pipette, and glassware calibration information, and CRM uncertainty. |
| p20252781 | W070 | Uncertainty estimate is based on method reproducibility based off of historic control data, pipette, and glassware calibration information, and CRM uncertainty. |
| p20252782 | W070 | Uncertainty estimate is based on method reproducibility based off of historic control data, pipette, and glassware calibration information, and CRM uncertainty. |
| p20252783 | W070 | Uncertainty estimate is based on method reproducibility based off of historic control data, pipette, and glassware calibration information, and CRM uncertainty. |

| UTIC | Webcode | Method for Estimating Uncertainty: |
|-----------|---------|---|
| p20252784 | W055 | Uncertainty estimate is based off of: pipette external calibration, pipette operator variance, gas chromatography variance, ethanol standard variance, volumetric flask variance, and volumetric pipette variance. |
| p20252785 | W055 | Uncertainty budget |
| p20252786 | W055 | Uncertainty estimated is based on method reproducibility, equipment variance, and operator variance. |
| p20252787 | W055 | Uncertainty estimate is based off of: pipette external calibration, pipette operator variance, gas chromatography variance, ethanol standard variance, volumetric flask variance, and volumetric pipette variance |
| p20252788 | W004 | Method uncertainty is estimated utilizing the method reproducibility (uncertainty of quality control historical data tracked by control charts), uncertainties for diluters, NIST SRM, and pipettes, and volumetric flask ATSM tolerance. |
| p20252789 | W037 | Uncertainty estimate is based on repeatability of positive controls; CRM uncertainty; replicate sample requirement and pipette uncertainty |

6) Item 1: Please report your quantitation result, reported in % (v/v):

Example: 17.1%.

7) Item 1: Please report your expanded measurement uncertainty, reported in %. If no uncertainty is reported "0" will be used for statistical purposes.

Example: 2.6 %.

8) Item 1: Please report the coverage factor (k) for your measurement uncertainty:

Example: k=2 (confidence level of approximately 95%, assuming the net mass follows a normal distribution)

| UTIC | Webcode | Item 1: Please report your quantitation result, reported in % (v/v) | Item 1: Please report your expanded measurement uncertainty, reported in % | Item 1: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252701 | W049 | 27.8% | 1.2% | k=2 (exp coverage probability of 95%) |
| p20252702 | W163 | 30.3 | 0.0 | (blank) |
| p20252703 | W088 | 28.1% | 1.4% | k = 2.58 (2.576) |
| p20252704 | W088 | 27.7% | 1.4% | 2.58 |
| p20252705 | W088 | 28.0% | 1.4% | k= 2.58 (2.578) |

| UTIC | Webcode | Item 1: Please report your quantitation result, reported in % (v/v) | Item 1: Please report your expanded measurement uncertainty, reported in % | Item 1: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252706 | W088 | 27.7% | 1.4% | k = 2.576 |
| p20252707 | W088 | 27.3% | 1.4% | 2.576 (99%) |
| p20252708 | W088 | 27.7% | 1.4% | k=2.58 (2.576) |
| p20252709 | W088 | 27.6% | 1.4% | k=2.58 (2.576) |
| p20252710 | W088 | 27.9% | 1.4% | k= 2.58 (2.576) |
| p20252711 | W088 | 28.3% | 1.4% | 2.58 |
| p20252712 | W088 | 28.1% | 1.4% | k=2.58 (2.576) |
| p20252713 | W088 | 27.5 | 1.4 | k=2.58 |
| p20252714 | W088 | 27.2% | 1.4% | k=2.576 |
| p20252715 | W088 | 27.9% | 1.4% | k=2.58 (2.576) |
| p20252716 | W088 | 26.9% | 1.3% | k=2.58 |
| p20252717 | W088 | 27.3 | 1.4 | k=2.58 |
| p20252718 | W088 | 27.4% | 1.4% | k= 2.58 (2.576) |
| p20252719 | W088 | 27.7% | 1.4% | 2.58 |
| p20252720 | W088 | 27.2 | 1.4 | k=2.576 |
| p20252721 | W088 | 27.1% | 1.4% | k = 2.576 |
| p20252722 | W088 | 28.5 | 1.4 | 2.58 |
| p20252723 | W088 | 27.8 | 1.4 | 2.58 |
| p20252724 | W088 | 27.8% | 1.4% | 2.58 |
| p20252725 | W088 | 27.5% | 1.4 | 2.58 |
| p20252726 | W088 | 27.5% | 1.4% | k = 2.58 |
| p20252727 | W088 | 27.9% | 1.4% | k = 2.58 |
| p20252728 | W088 | 28.3% | 1.4% | 2.58 |
| p20252733 | W092 | 27.2% | 1.1% | k=2 |
| p20252735 | W031 | 26.8667% | 2.5059% | k=3 |
| p20252736 | W079 | Greater than 1/2 of 1% by volume | 0 | k=2 |
| p20252737 | W079 | Greater than 1/2 of 1% by volume | 0 | k=2 |
| p20252738 | W044 | 26.3% | 3.4% | 3 |
| p20252739 | W114 | 27.3% | 2.2% | k=3 (Confidence level 99.73%) |
| p20252740 | W114 | 27.6% | 2.2% | k = 3 ; Confidence level 99.73% |
| p20252741 | W114 | 29.1% | 2.3% | k=3 (confidence level 99.73%) |
| p20252742 | W114 | 28.0% | 2.2% | k = 3 ; Confidence level 99.73% |
| p20252743 | W114 | 28.7% | 2.3% | Coverage factor k=3; Conf. Level 99.73% |

| UTIC | Webcode | Item 1: Please report your quantitation result, reported in % (v/v) | Item 1: Please report your expanded measurement uncertainty, reported in % | Item 1: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252744 | W114 | 27.3% | +/- 2.2% | k = 3 (Confidence level of 99.73%) |
| p20252745 | W114 | 27.8% | 2.2% | k=3 (Confidence Level 99.73 %) |
| p20252746 | W114 | 27.3% | 2.2% | 3 |
| p20252747 | W114 | 27.7% | 2.2% | k=3 (confidence level 99.73%) |
| p20252749 | W190 | 27.7% | 2.2% | k=3, Confidence level 99.73% |
| p20252750 | W190 | 27.1% | 2.2% | k= 3 confidence level: 99.73 % |
| p20252751 | W190 | 27.7% | 2.2% | k=3 (confidence level of 99.73%) |
| p20252752 | W190 | 27.3% | 2.2% | k=3; Confidence level 99.73% |
| p20252753 | W190 | 28.2% | 2.2% | k=3; confidence level 99.73% |
| p20252754 | W190 | 27.2% | 2.2% | k=3; Confidence level 99.73% |
| p20252755 | W190 | 27.1% | 2.2% | Coverage factor k = 3; Confidence 99.73% |
| p20252756 | W190 | 27.3% | 2.2% | k=3 (confidence level 99.73%) |
| p20252757 | W190 | 27.2% | 2.2% | k = 3; Confidence level 99.73% |
| p20252758 | W002 | 29.5% | 1.6% | k=3 |
| p20252761 | W110 | 29.6% | +/- 2.1% | K = 3 (confidence level of ~ 99%) |
| p20252762 | W110 | 28.0% | ± 2.0% | k=3 (confidence level of approx 99%) |
| p20252766 | W110 | 27.9% | 2.0% | k=3 |
| p20252769 | W110 | 26.5% | +/- 1.9% | K=3 (confidence level of ~ 99%) |
| p20252770 | W082 | 27.5% | 2.0% | k=2 (~95% level of confidence) |
| p20252771 | W172 | 27 | 3 | 2 |
| p20252772 | W096 | 27.8% | 1.0% | 2.05 |
| p20252775 | W206 | 26.3% | 2.1% | k=2 |
| p20252776 | W116 | 27.5% | 2.6% | k=3 |
| p20252777 | W116 | 28.3% | 2.7% | k=3 |
| p20252778 | W025 | 26.7% | 0.3% | k=3 |
| p20252779 | W256 | 28.2% (v/v) | Not Applicable | Not Applicable |
| p20252780 | W070 | 28.1% | 1.4% | k = 3 (confidence level of 99.73%) |
| p20252781 | W070 | 27.6% | 1.3% | K=3 (confidence level of 99.73%) |

| UTIC | Webcode | Item 1: Please report your quantitation result, reported in % (v/v) | Item 1: Please report your expanded measurement uncertainty, reported in % | Item 1: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252782 | W070 | 28.2% | 1.4% | K=3 (confidence level of 99.73%) |
| p20252783 | W070 | 27.4% | 1.3% | K=3 (confidence level of 99.73%) |
| p20252784 | W055 | 28.7% | 1.9% | k=2, 95% |
| p20252785 | W055 | 27.4% | 1.8% | 2 |
| p20252786 | W055 | 29.1% | 1.9% | 2 |
| p20252787 | W055 | 28.7 | 1.9 | k=2. 95% |
| p20252788 | W004 | 28% | 1% | k=2 (95.45 % confidence level) |
| p20252789 | W037 | 27.94% | 1.63% | 3 |

9) Item 2: Please report your quantitation result, reported in % (v/v):

Example: 17.1%.

10) Item 2: Please report your expanded measurement uncertainty, reported in %. If no uncertainty is reported "0" will be used for statistical purposes.

Example: 2.6 %.

11) Item 2: Please report the coverage factor (k) for your measurement uncertainty:

Example: k=2 (confidence level of approximately 95%, assuming the net mass follows a normal distribution)

| UTIC | Webcode | Item 2: Please report your quantitation result, reported in % (v/v) | Item 2: Please report your expanded measurement uncertainty, reported in % | Item 2: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252701 | W049 | 5.0% | 0.6% | k=2 (exp coverage probability of 95%) |
| p20252702 | W163 | 5.1 | 0.0 | (blank) |
| p20252703 | W088 | 5.09% | 0.25% | k = 2.58 (2.576) |
| p20252704 | W088 | 5.01% | 0.25% | 2.58 |

| UTIC | Webcode | Item 2: Please report your quantitation result, reported in % (v/v) | Item 2: Please report your expanded measurement uncertainty, reported in % | Item 2: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252705 | W088 | 5.00% | 0.25% | k= 2.58 (2.578) |
| p20252706 | W088 | 4.99% | 0.25% | k = 2.576 |
| p20252707 | W088 | 5.09% | 0.25% | 2.576 (99%) |
| p20252708 | W088 | 5.65% | 0.28% | k=2.58 (2.576) |
| p20252709 | W088 | 4.89% | 0.24% | k=2.58 (2.576) |
| p20252710 | W088 | 5.04% | 0.25% | k= 2.58 (2.576) |
| p20252711 | W088 | 5.08% | 0.25% | 2.58 |
| p20252712 | W088 | 4.95% | 0.25% | k=2.58 (2.576) |
| p20252713 | W088 | 4.95 | 0.25 | k=2.58 |
| p20252714 | W088 | 4.85% | 0.24% | k=2.576 |
| p20252715 | W088 | 4.98% | 0.25% | k=2.58 (2.576) |
| p20252716 | W088 | 4.85% | 0.24% | k=2.58 |
| p20252717 | W088 | 4.94 | 0.25 | k=2.58 |
| p20252718 | W088 | 4.89% | 0.24% | k= 2.58 (2.576) |
| p20252719 | W088 | 4.95% | 0.25% | 2.58 |
| p20252720 | W088 | 4.90 | 0.25 | k=2.576 |
| p20252721 | W088 | 4.63% | 0.23% | k = 2.576 |
| p20252722 | W088 | 5.13 | 0.26 | 2.58 |
| p20252723 | W088 | 5.05 | 0.25 | 2.58 |
| p20252724 | W088 | 5.00% | 0.25% | 2.58 |
| p20252725 | W088 | 5.06% | 0.25 | 2.58 |
| p20252726 | W088 | 4.98% | 0.25% | k = 2.58 |
| p20252727 | W088 | 5.07% | 0.25% | k = 2.58 |
| p20252728 | W088 | 5.16% | 0.26% | 2.58 |
| p20252733 | W092 | 5.0% | 0.2% | k=2 |
| p20252735 | W031 | 4.8915% | 0.4562% | k=3 |
| p20252736 | W079 | Greater than 1/2 of 1% by volume | 0 | k=2 |
| p20252737 | W079 | Greater than 1/2 of 1% by volume | 0 | k=2 |
| p20252738 | W044 | 4.97% | 0.65% | 3 |
| p20252739 | W114 | 5.0% | 0.5% | k=3 (Confidence level 99.73%) |
| p20252740 | W114 | 4.9% | 0.4% | k = 3 ; Confidence level 99.73% |
| p20252741 | W114 | 4.9% | 0.4% | k=3 (confidence level 99.73%) |

| UTIC | Webcode | Item 2: Please report your quantitation result, reported in % (v/v) | Item 2: Please report your expanded measurement uncertainty, reported in % | Item 2: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252742 | W114 | 4.9% | 0.4% | k = 3 ; Confidence level 99.73% |
| p20252743 | W114 | 4.9% | 0.4% | Coverage factor k=3; Conf. Level 99.73% |
| p20252744 | W114 | 4.8% | +/- 0.4% | k = 3 (Confidence level of 99.73%) |
| p20252745 | W114 | 4.8% | 0.4% | k=3 (Confidence Level 99.73 %) |
| p20252746 | W114 | 5.0% | 0.5% | 3 |
| p20252747 | W114 | 4.8% | 0.4% | k=3 (confidence level 99.73%) |
| p20252749 | W190 | 4.9% | 0.4% | k=3, Confidence level 99.73% |
| p20252750 | W190 | 4.9% | 0.4% | k= 3 confidence level: 99.73 % |
| p20252751 | W190 | 5.1% | 0.5% | k=3 (confidence level of 99.73%) |
| p20252752 | W190 | 4.9% | 0.4% | k=3; Confidence level 99.73% |
| p20252753 | W190 | 5.0% | 0.6% | k=3; confidence level 99.73 % |
| p20252754 | W190 | 4.9% | 0.4% | k=3; Confidence level 99.73% |
| p20252755 | W190 | 4.8% | 0.4% | Coverage factor k = 3; Confidence 99.73% |
| p20252756 | W190 | 4.8% | 0.4% | k=3 (confidence level 99.73%) |
| p20252757 | W190 | 5.0% | 0.5% | k = 3; Confidence level 99.73% |
| p20252758 | W002 | 4.9% | 0.3% | k=3 |
| p20252761 | W110 | 5.2% | +/- 0.4% | K = 3 (confidence level of ~ 99%) |
| p20252762 | W110 | 5.0% | ± 0.4%. | k=3 (confidence level of approx 99%) |
| p20252766 | W110 | 5.4% | 0.4% | k=3 |
| p20252769 | W110 | 4.8% | +/- 0.4% | K=3 (confidence level of ~ 99%) |
| p20252770 | W082 | 5.1% | 0.4% | k=2 (~95% level of confidence) |
| p20252771 | W172 | 4.7 | 0.5 | 2 |
| p20252772 | W096 | 5.05% | 0.17% | 2.05 |
| p20252775 | W206 | 4.7% | 0.3% | k=2 |

| UTIC | Webcode | Item 2: Please report your quantitation result, reported in % (v/v) | Item 2: Please report your expanded measurement uncertainty, reported in % | Item 2: Please report the coverage factor (k) for your measurement uncertainty |
|-----------|---------|---|--|--|
| p20252776 | W116 | 5.1% | 0.5% | k=3 |
| p20252777 | W116 | 4.9% | 0.5% | k=3 |
| p20252778 | W025 | 4.7% | 0.1% | k=3 |
| p20252779 | W256 | 5.0% (v/v) | Not Applicable | Not Applicable |
| p20252780 | W070 | 5.1% | 0.3% | k = 3 (confidence level of 99.73%) |
| p20252781 | W070 | 5.1% | 0.2% | K=3 (confidence level of 99.73%) |
| p20252782 | W070 | 5.0% | 0.3% | K=3 (confidence level of 99.73%) |
| p20252783 | W070 | 5.0% | 0.2% | K=3 (confidence level of 99.73%) |
| p20252784 | W055 | 4.99% | 0.33% | k=2, 95% |
| p20252785 | W055 | 5.01% | 0.33% | 2 |
| p20252786 | W055 | 4.88% | 0.32% | 2 |
| p20252787 | W055 | 5.03 | 0.33 | k=2, 95% |
| p20252788 | W004 | 5.1% | 0.1% | k=2 (95.45% confidence level) |
| p20252789 | W037 | 5.19% | 0.30% | 3 |

12) How long did it take to complete this test (in hours)? Please report actual analytical hours only.

13) Did you find this test to be a fair test of the process of the quantitation of ethanol?

A) Yes
 B) No

| UTIC | Webcode | How long did it take to complete this test (in hours)? Please report actual analytical hours only. | Did you find this test to be a fair test of the process of the quantitation of ethanol? |
|-----------|---------|--|---|
| p20252701 | W049 | 8 hours - entire run (w/ other samples) | Yes |
| p20252702 | W163 | (blank) | Yes |
| p20252703 | W088 | 3 | Yes |

| UTIC | Webcode | How long did it take to complete this test (in hours)? Please report actual analytical hours only. | Did you find this test to be a fair test of the process of the quantitation of ethanol? |
|-----------|---------|--|---|
| p20252704 | W088 | 1 | Yes |
| p20252705 | W088 | 3 | Yes |
| p20252706 | W088 | 1 | Yes |
| p20252707 | W088 | 2.5 hours | Yes |
| p20252708 | W088 | less than 2 hours | Yes |
| p20252709 | W088 | 2 | Yes |
| p20252710 | W088 | 2.5 | Yes |
| p20252711 | W088 | ~ 3 hrs | Yes |
| p20252712 | W088 | 3 | Yes |
| p20252713 | W088 | 1 | Yes |
| p20252714 | W088 | 1 hour | Yes |
| p20252715 | W088 | 1 | Yes |
| p20252716 | W088 | 0.25 | Yes |
| p20252717 | W088 | 0.25 | Yes |
| p20252718 | W088 | 1 | Yes |
| p20252719 | W088 | 1 | Yes |
| p20252720 | W088 | 1.5 | Yes |
| p20252721 | W088 | 0.5 | Yes |
| p20252722 | W088 | 1.5 | Yes |
| p20252723 | W088 | 1.5 | Yes |
| p20252724 | W088 | 1.5 | Yes |
| p20252725 | W088 | 2 | Yes |
| p20252726 | W088 | 1 hr | Yes |
| p20252727 | W088 | 1 | Yes |
| p20252728 | W088 | 1 hour | Yes |
| p20252733 | W092 | 3 hours | Yes |
| p20252735 | W031 | approximately 2 hours | Yes |
| p20252736 | W079 | 3 | Yes |
| p20252737 | W079 | 2 | Yes |
| p20252738 | W044 | 2.5 | Yes |
| p20252739 | W114 | 4.5 hours | Yes |
| p20252740 | W114 | 2.5 hrs | Yes |
| p20252741 | W114 | 2 hours | Yes |
| p20252742 | W114 | 2 hours | Yes |
| p20252743 | W114 | 2 | Yes |
| p20252744 | W114 | 4 hours | Yes |
| p20252745 | W114 | 1 | Yes |
| p20252746 | W114 | 2 | Yes |
| p20252747 | W114 | 1 hour | Yes |
| p20252749 | W190 | 1 | Yes |

| UTIC | Webcode | How long did it take to complete this test (in hours)? Please report actual analytical hours only. | Did you find this test to be a fair test of the process of the quantitation of ethanol? |
|-----------|---------|--|---|
| p20252750 | W190 | 1 | Yes |
| p20252751 | W190 | 2 | Yes |
| p20252752 | W190 | 3 | Yes |
| p20252753 | W190 | 1 hour | Yes |
| p20252754 | W190 | approximately 2 hours | Yes |
| p20252755 | W190 | 2 | Yes |
| p20252756 | W190 | 1.5 | Yes |
| p20252757 | W190 | 2.5 | Yes |
| p20252758 | W002 | 2hrs | Yes |
| p20252761 | W110 | ~3.0 hours | Yes |
| p20252762 | W110 | 3 | Yes |
| p20252766 | W110 | 3 | Yes |
| p20252769 | W110 | 1.5 hours | Yes |
| p20252770 | W082 | 3 hours | Yes |
| p20252771 | W172 | 4 | Yes |
| p20252772 | W096 | 8 | Yes |
| p20252775 | W206 | 5 Hours | Yes |
| p20252776 | W116 | 2 | Yes |
| p20252777 | W116 | 2 | Yes |
| p20252778 | W025 | 3 | Yes |
| p20252779 | W256 | 2 | Yes |
| p20252780 | W070 | 3 hrs | Yes |
| p20252781 | W070 | 2.5 hours | Yes |
| p20252782 | W070 | 2.5 hours | Yes |
| p20252783 | W070 | 2.5 hours | Yes |
| p20252784 | W055 | 6 | Yes |
| p20252785 | W055 | 5 | Yes |
| p20252786 | W055 | 8 | Yes |
| p20252787 | W055 | 4 | Yes |
| p20252788 | W004 | 3-4 hours | Yes |
| p20252789 | W037 | 6 | Yes |

14) How would you change the aspects of the test (i.e. scenario, test samples, question sections, report format) to improve a future version of this test? Comments and suggestions are welcome.

Additionally, this question is a means to provide you with an opportunity to explain or include information about your findings or interpretation, as needed. In order to maintain confidentiality, please refrain from including identifying information specific to your laboratory.

| UTIC | Webcode | How would you change the aspects of the test (i.e. scenario, test samples, question sections, report format) to improve a future version of this test? Comments and suggestions are welcome. | FTS Response |
|-----------|---------|---|--|
| p20252701 | W049 | Test was well put together, data entry is simple. I'm still confused why so many method specifics are requested like the column and standard info. | Thank you for the comment. FTS requests information regarding methods, standards, and uncertainties to provide additional context in our Summary Report for other agencies to utilize and learn. |
| p20252716 | W088 | Needs clarification on question 4 (which standard, what's used to build the calibration curve, what's used with the controls?). Spell check doesn't work in Google Chrome. Would like clarification on sig figs for answers (2 after the decimal? 3 total sig figs?). | Thank you for the suggestion. We ask that you report as you would for normal casework. FTS will adjust the significant figures, if necessary, for statistical purposes. |
| p20252735 | W031 | I found the overall test samples and questions to be straightforward. Submitted evidence was stored in the refrigerator upon receipt. Calibrators, controls, and evidence were removed from the refrigerator and allowed to warm to room temperature for at least thirty minutes before sampling and diluting. | |

| UTIC | Webcode | How would you change the aspects of the test (i.e. scenario, test samples, question sections, report format) to improve a future version of this test? Comments and suggestions are welcome. | FTS Response |
|-----------|---------|--|--|
| p20252738 | W044 | <p>I recommend the use of evidence tape or tape that is easier to tear/open with gloves on (in reference to the vial itself).</p> <p>Our results are normally reported in g/100 mL and had to be converted for this PT. The following reflects our data before using ethanol's density (0.789 g/mL) for the conversion:</p> <p>Item 1 results: 20.7 g/100 mL Item 1 UM: 2.7 g/100 mL</p> <p>Item 2 results: 3.92 g/100 mL Item 2 UM: 0.51 g/100 mL</p> | <p>Thank you for the suggestion.</p> <p>Thank you for clarifying your results.</p> |
| p20252741 | W114 | Additional information in the form of a case scenario might be helpful for choosing what volume glassware to use. | Thank you for the suggestion. |
| p20252744 | W114 | Use of narratives to inform initial sample dilution. This could also be accomplished with using commercially available alcohol products so that scent wafting could inform sample dilution. | Thank you for the suggestion. We will consider this for future PTs. |
| p20252745 | W114 | For dilution purposes, it would be helpful to have a description of the sample: Sample retrieved from a bottle labeled "Miller Lite." | Thank you for the suggestion. We will consider this for future PTs. |
| p20252747 | W114 | I think using juice as the matrix for one of the unknowns would be helpful and better represent an actual sample received at the laboratory. | Thank you for the suggestion. |
| p20252755 | W190 | None, good test. | |
| p20252756 | W190 | Utilizing a different matrix to better simulate a real-world scenario | Thank you for the suggestion. |

| UTIC | Webcode | How would you change the aspects of the test (i.e. scenario, test samples, question sections, report format) to improve a future version of this test? Comments and suggestions are welcome. | FTS Response |
|-----------|---------|---|---|
| p20252758 | W002 | Maybe try something different than a clear liquid net time. | Thank you for the suggestion. |
| p20252770 | W082 | This test could do with more test samples, including a negative sample. | Thank you for the suggestion. |
| p20252771 | W172 | question sections are unnecessary | Please see FTS Response for p20252701. |
| p20252776 | W116 | The question section is more in-depth than other proficiency test providers. Would recommend solutions that are different colors and clearness to better represent casework samples. | Please see FTS Response for p20252701. Thank you for the suggestion. |
| p20252777 | W116 | The question section is more in-depth than other proficiency test providers. Would recommend solutions that are different colors and clearness to better represent casework samples. | Please see FTS Response for p20252701. Thank you for the suggestion. |
| p20252778 | W025 | An ounce of sample provided instead of only half an ounce. | Thank you for the suggestion. FTS will consider this for future PTs. |
| p20252786 | W055 | Include more realistic beverages into the mix. It seems like you could buy a commercial whiskey and wine in single container and sufficient volume to be fit for purpose for distribution and validation needs. | Thank you for the suggestion. We will consider this for future PTs. |